

Freedom of Information Request	FOI 21-249	2 nd July 2021
--------------------------------	------------	---------------------------

I request that you inform me of the number of PCR test cycles you use to detect the presence of COVID-19 VIRUS.

The Health Board uses several platforms for testing, these are Cepheid, Eplex, Biofire and Seegene. They vary in cycle numbers but the average is 45.