

Freedom of Information Request	FOI 21-253	5 th July 2021
--------------------------------	------------	---------------------------

1: How many cycles are your PCR tests run at?

The Health Board uses several platforms for testing, these are Cepheid, Eplex, Biofire and Seegene. They vary in cycle numbers but the average is 45.

2: Have you ever altered the cycles for a period of time, if yes, at what Amplification?

No